EC-Certification of Conformity

Certification of Conformity

EU-MEDICAL DEVICES DIRECTIVE-93/42/EEC

Registration No.: ENC180106GZ35

Applicant	:	Guangzhou Linuo Beauty Electronic Technology Co., Ltd.
Address	:	Room 215, No.7, Huangshagang Street, Xinshi Street, Baiyun District, Guangzhou City, China
Product Designation	:	Derma roller
Model Number	:	MT-540, MT-T540, MT-75, MT-180, MT-192, MT-200, MT-240, MT-600, MT-1080, MT-1200, LN-S540, LN-T540, LN-31, LN-4 IN 1, LN-5 IN 1, LN-16R1, LN-S1, LN-S2, LN-S3, LN-S4, LN-S5, ZGTS-540, ZGTS-T540, ZGTS-192, DRS-40, DRS-192, DRS-540, JMF-540, JMF-75, JMF-180, JMF-192, JMF-200, JMF-240, JMF-600, JMF-1080, JMF-1200, ZT-12, ZT-36, DNS-192, MNS-192

Classification of Medical Devices: Class I

Manufacturer	:	Guangzhou Linuo Beauty Electronic Technology Co., Ltd.
Address	:	Room 215, No.7, Huangshagang Street, Xinshi Street, Baiyun District, Guangzhou City, China

The submitted products have been tested by us with the listed standard and found in compliance with the Medical Devices Directive 93/42/EEC and following European Standards:

Requirement	Applied Standards	Test Report No.
DELAYED-TYPE HYPERSENSIVITY	EN ISO 10993-1:2009	ENC180106GZ35E1
	EN ISO 10993-10:2013	
	EN ISO 10993-12:2012	

Test result: Derma roller did not cause skin irritation or allergic sensitization.

This certificate is based on an evaluation of a sample of the above mentioned product. Technical report and documentation are at the licence applicant's disposal. This certificate does not imply assessment of the series-production of the product. The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

CE

East Notice Certification Service Co., Ltd. ENC

Ray Zhou / General Manager

Date of Issue: Jan. 6, 2018

East Notice Certification Service Co., Ltd.

1/F, Haohui Commercial Building, Zhuji Street, Dongpu Town, Tianhe District, Guangzhou City

Tel:+86-020-2331 4234 Fax:+86-020-8256 8534 E-mail:enc@enc-lab.com Web: www.enc-lab.com